

High-Volume TSP Sampler
5-Point Calibration Record

Location : ASR 5
 Calibrated by : P.F.Yeung
 Date : 09/10/2013

Sampler

Model : TE-5170
 Serial Number : S/N 0816

Calibration Office and Standard Calibration Relationship

Serial Number : 2323
 Service Date : 26 Dec 2012
 Slope (m) : 2.09107
 Intercept (b) : -0.02838
 Correlation Coefficient(r) : 0.99996

Standard Condition

Pstd (hpa) : 1013
 Tstd (K) : 298.18

Calibration Condition

Pa (hpa) : 1017
 Ta(K) : 299

Resistance Plate	dH [green liquid] (inch water)	Z	X=Qstd (cubic meter/min)	IC (chart)	Y (corrected)
1 18 holes	12.5	3.537	1.705	52	52.02
2 13 holes	9.7	3.115	1.503	45	45.01
3 10 holes	7.6	2.758	1.332	40	40.01
4 7 holes	4.7	2.169	1.051	31	31.01
5 5 holes	3.0	1.733	0.842	24	24.01

Notes: $Z = \sqrt{dH(Pa/Pstd)(Tstd/Ta)}$, $X = Z/m - b$, $Y(\text{Corrected Flow}) = IC * \{\sqrt{Pa/Pstd}(Tstd/Ta)\}$

Sampler Calibration Relationship (Linear Regression)

Slope(m): 32.148 Intercept(b): -2.953 Correlation Coefficient(r): 0.9997

Checked by: Magnum Fan

Date: 20/10/2013

High-Volume TSP Sampler
5-Point Calibration Record

Location : ASR10A
 Calibrated by : P.F.Yeung
 Date : 15/10/2013

Sampler

Model : TE-5170
 Serial Number : S/N 8162

Calibration Office and Standard Calibration Relationship

Serial Number : 2323
 Service Date : 26 Dec 2012
 Slope (m) : 2.09107
 Intercept (b) : -0.02838
 Correlation Coefficient(r) : 0.99996

Standard Condition

Pstd (hpa) : 1013
 Tstd (K) : 298.18

Calibration Condition

Pa (hpa) : 1013
 Ta(K) : 301

Resistance Plate	dH [green liquid] (inch water)	Z	X=Qstd (cubic meter/min)	IC (chart)	Y (corrected)
1 18 holes	13.0	3.588	1.729	59	58.71
2 13 holes	10.4	3.209	1.548	52	51.74
3 10 holes	7.8	2.779	1.343	45	44.78
4 7 holes	5.0	2.225	1.078	36	35.82
5 5 holes	3.0	1.723	0.838	28	27.86

Notes: $Z = \sqrt{dH(Pa/Pstd)(Tstd/Ta)}$, $X = Z/m - b$, $Y(\text{Corrected Flow}) = IC * \{\sqrt{Pa/Pstd}(Tstd/Ta)\}$

Sampler Calibration Relationship (Linear Regression)

Slope(m): 34.384 Intercept(b): 1.161 Correlation Coefficient(r): 0.9997

Checked by: Magnum Fan

Date: 20/10/2013

High-Volume TSP Sampler
5-Point Calibration Record

Location : AQM1
 Calibrated by : P.F.Yeung
 Date : 17/10/2013

Sampler

Model : TE-5170
 Serial Number : S/N 1253

Calibration Office and Standard Calibration Relationship

Serial Number : 2323
 Service Date : 26 Dec 2012
 Slope (m) : 2.09107
 Intercept (b) : -0.02838
 Correlation Coefficient(r) : 0.99996

Standard Condition

Pstd (hpa) : 1013
 Tstd (K) : 298.18

Calibration Condition

Pa (hpa) : 1017
 Ta(K) : 299

Resistance Plate	dH [green liquid] (inch water)	Z	X=Qstd (cubic meter/min)	IC (chart)	Y (corrected)
1 18 holes	13.4	3.662	1.765	56	56.02
2 13 holes	9.4	3.067	1.480	47	47.01
3 10 holes	7.5	2.739	1.324	41	41.01
4 7 holes	5.0	2.237	1.083	33	33.01
5 5 holes	3.0	1.733	0.842	26	26.01

Notes: $Z = \sqrt{dH(Pa/Pstd)(Tstd/Ta)}$, $X = Z/m - b$, $Y(\text{Corrected Flow}) = IC * \{\sqrt{dH(Pa/Pstd)(Tstd/Ta)}\}$

Sampler Calibration Relationship (Linear Regression)

Slope(m): 32.944 Intercept(b): -2.175 Correlation Coefficient(r): 0.9990

Checked by: Magnum Fan

Date: 20/10/2013

High-Volume TSP Sampler
5-Point Calibration Record

Location : ASR 1
 Calibrated by : P.F.Yeung
 Date : 17/10/2013

Sampler

Model : TE-5170
 Serial Number : S/N 0146

Calibration Office and Standard Calibration Relationship

Serial Number : 2323
 Service Date : 26 Dec 2012
 Slope (m) : 2.09107
 Intercept (b) : -0.02838
 Correlation Coefficient(r) : 0.99996

Standard Condition

Pstd (hpa) : 1013
 Tstd (K) : 298.18

Calibration Condition

Pa (hpa) : 1016
 Ta(K) : 299

Resistance Plate	dH [green liquid] (inch water)	Z	X=Qstd (cubic meter/min)	IC (chart)	Y (corrected)
1 18 holes	13.0	3.605	1.737	52	51.99
2 13 holes	10.4	3.224	1.555	46	45.99
3 10 holes	7.8	2.792	1.349	39	38.99
4 7 holes	5.0	2.236	1.083	30	29.99
5 5 holes	3.0	1.732	0.842	23	22.99

Notes: $Z = \sqrt{dH(Pa/Pstd)(Tstd/Ta)}$, $X = Z/m - b$, $Y(\text{Corrected Flow}) = IC * \{\sqrt{Pa/Pstd}(Tstd/Ta)\}$

Sampler Calibration Relationship (Linear Regression)

Slope(m): 32.647 Intercept(b): -4.881 Correlation Coefficient(r): 0.9996

Checked by: Magnum Fan

Date: 20/10/2013

High-Volume TSP Sampler
5-Point Calibration Record

Location : ASR 6A
 Calibrated by : P.F.Yeung
 Date : 17/10/2013

Sampler

Model : TE-5170
 Serial Number : S/N 1059

Calibration Office and Standard Calibration Relationship

Serial Number : 2323
 Service Date : 26 Dec 2012
 Slope (m) : 2.09107
 Intercept (b) : -0.02838
 Correlation Coefficient(r) : 0.99996

Standard Condition

Pstd (hpa) : 1013
 Tstd (K) : 298.18

Calibration Condition

Pa (hpa) : 1017
 Ta(K) : 299

Resistance Plate	dH [green liquid] (inch water)	Z	X=Qstd (cubic meter/min)	IC (chart)	Y (corrected)
1 18 holes	12.5	3.537	1.705	56	56.02
2 13 holes	10.0	3.163	1.526	50	50.01
3 10 holes	8.0	2.829	1.367	44	44.01
4 7 holes	5.2	2.281	1.104	35	35.01
5 5 holes	2.8	1.674	0.814	26	26.01

Notes: $Z = \sqrt{dH(Pa/Pstd)(Tstd/Ta)}$, $X = Z/m - b$, $Y(\text{Corrected Flow}) = IC * \{\sqrt{Pa/Pstd}(Tstd/Ta)\}$

Sampler Calibration Relationship (Linear Regression)

Slope(m): 32.148 Intercept(b): -2.953 Correlation Coefficient(r): 0.9997

Checked by: Magnum Fan

Date: 20/10/2013

High-Volume TSP Sampler
5-Point Calibration Record

Location : ASR 9C
 Calibrated by : P.F.Yeung
 Date : 05/11/2013

Sampler

Model : TE-5170
 Serial Number : S/N 3572

Calibration Office and Standard Calibration Relationship

Serial Number : 2323
 Service Date : 26 Dec 2012
 Slope (m) : 2.09107
 Intercept (b) : -0.02838
 Correlation Coefficient(r) : 0.99996

Standard Condition

Pstd (hpa) : 1013
 Tstd (K) : 298.18

Calibration Condition

Pa (hpa) : 1019
 Ta(K) : 296

Resistance Plate	dH [green liquid] (inch water)	Z	X=Qstd (cubic meter/min)	IC (chart)	Y (corrected)
1 18 holes	11.4	3.398	1.638	56	56.36
2 13 holes	9.0	3.019	1.457	50	50.32
3 10 holes	6.6	2.585	1.250	43	43.27
4 7 holes	4.2	2.062	0.100	36	36.23
5 5 holes	2.4	1.559	0.759	29	29.18

Notes: $Z = \sqrt{dH(Pa/Pstd)(Tstd/Ta)}$, $X = Z/m - b$, $Y(\text{Corrected Flow}) = IC * \{\sqrt{Pa/Pstd}(Tstd/Ta)\}$

Sampler Calibration Relationship (Linear Regression)

Slope(m): 30.801 Intercept(b): 5.465 Correlation Coefficient(r): 0.9991

Checked by: Magnum Fan

Date: 08/11/2013

High-Volume TSP Sampler
5-Point Calibration Record

Location : ASR9A
 Calibrated by : P.F.Yeung
 Date : 05/11/2013

Sampler

Model : TE-5170
 Serial Number : S/N 3573

Calibration Office and Standard Calibration Relationship

Serial Number : 2323
 Service Date : 26 Dec 2012
 Slope (m) : 2.09107
 Intercept (b) : -0.02838
 Correlation Coefficient(r) : 0.99996

Standard Condition

Pstd (hpa) : 1013
 Tstd (K) : 298.18

Calibration Condition

Pa (hpa) : 1019
 Ta(K) : 296

Resistance Plate	dH [green liquid] (inch water)	Z	X=Qstd (cubic meter/min)	IC (chart)	Y (corrected)
1 18 holes	10.5	3.261	1.573	60	60.38
2 13 holes	8.2	2.882	1.392	54	54.34
3 10 holes	6.0	2.465	1.192	49	49.31
4 7 holes	4.0	2.013	0.976	42	42.27
5 5 holes	2.6	1.623	0.790	37	37.23

Notes: $Z = \sqrt{dH(Pa/Pstd)(Tstd/Ta)}$, $X = Z/m - b$, $Y(\text{Corrected Flow}) = IC * \{\sqrt{Pa/Pstd}(Tstd/Ta)\}$

Sampler Calibration Relationship (Linear Regression)

Slope(m): 29.442 Intercept(b): 13.831 Correlation Coefficient(r): 0.9992

Checked by: Magnum Fan

Date: 08/11/2013



METHOD REFERENCE

Total Suspended Particulates

ALS Method Code: HK-TSP

New glass fiber filters are conditioned in a desiccator of RH <50% for 24 hours or more before use. The initial filter weight was taken and record in the record book. Then the filter was put in a clean envelope for transportation.

After sampling, the filter was put in a paper holder and condition in a desiccator of RH <50% for 24 hours or more. The filter was removed from the holder carefully before weighing. The final weight was taken and record in the record book.

The dust weight was calculated as the difference of the weight of filter before and after sampling.

WATER

Certification of Quality

This product has been tested in accordance with procedures established through Global Water Instrumentation's Quality Management System. This product meets or exceeds its manufacturing acceptance criteria.

ITEM DESCRIPTION:	Wind Speed Sensor
MODEL NAME/ NUMBER:	WE550
PART NUMBER:	EC0000
SENSOR RANGE:	0-110 MPH
SENSOR OUTPUT:	4.00-19.91 mA
ACCURACY:	.2 MPH over the range 11 to 55 MPH
POWER REQUIRED	10-36 VDC
SERIAL NUMBER:	1337005099
CABLE LENGTH:	25 ft
CERTIFICATES:	CE Compliant

Contact
Global Water
for all your
instrumentation
needs:

Water Level

Water Flow

Water Samplers

Water Quality

Weather

Remote Monitoring

Control

Technician: *Wright, Jess*

Date: 9/10/2013

NOT Global Water Instrumentation warrants that its products are free from defects in material & workmanship under normal use & service for a period of one year from date of original shipment from factory. Repaired components are warranted for a period of 90 days from shipment. Contact us for complete warranty details.



Global Water

a **xylem** brand

In the U.S. call toll free
at 1-800-876-1172

International: 1-979-690-5560

Fax: 1-979-690-0440

Email: globalw@globalw.com

Visit our online catalog at:
www.globalw.com

Our Service Address

151 Graham Rd

College Station, TX 77845

Certification of Quality

This product has been tested in accordance with procedures established through Global Water Instrumentation's Quality Management System. This product meets or exceeds its manufacturing acceptance criteria.

ITEM DESCRIPTION:	Wind Direction
MODEL NAME/ NUMBER:	WE570
PART NUMBER:	ED0000
SENSOR RANGE:	0-360 °
SENSOR OUTPUT:	4.01-20.03 mA
ACCURACY:	1% of full scale
POWER REQUIRED	10-36 VDC
SERIAL NUMBER:	1337005143
CABLE LENGTH:	25 ft
CERTIFICATES:	CE Compliant

Contact
Global Water
for all your
instrumentation
needs:

Water Level

Water Flow

Water Samplers

Water Quality

Weather

Remote Monitoring

Control

Technician: *Wright, Jess*

Date: 9/12/2013

NOT Global Water Instrumentation warrants that its products are free from defects in material & workmanship under normal use & service for a period of one year from date of original shipment from factory. Repaired components are warranted for a period of 90 days from shipment. Contact us for complete warranty details.



Global Water

a xylem brand

In the U.S. call toll free
at 1-800-878-1172
International: 1-979-690-5560
Fax: 1-979-690-0440
Email: globalw@globalw.com

Visit our online catalog at
www.globalw.com
Our Service Address
151 Graham Rd
College Station, TX 77845

Product Test Report



Product Tested: MetPak
Part Number: 1723-1B-2-111
Serial Number: 13130002
Test Date: 26/03/2013
Location: Gill Instruments Ltd

GILL ensures that quality is inherent in all aspects of their activities and ensures that compliance with BS EN ISO9001: 2008 is maintained.

This report certifies that the above instrument has been tested in accordance with Gill internal procedures

Results

Test	Limits	Results
Wind Still Air Test (Zero Wind Speed)	Pass/Fail	Pass
Wind Tunnel Test (12m/s nominal)	Pass/Fail	Pass
Pressure Sensor (Comparison DPI 142)	Pass/Fail	Pass
Temperature Sensor (Comparison HC2-S (SCS certified))	Pass/Fail	Pass
Humidity Sensor (Comparison HC2-S (SCS certified))	Pass/Fail	Pass

Wind sensor generic calibration is traceable to the University of Southampton wind tunnel and Gill instrumentation is maintained in accordance with UKAS.

Comparisons for Temperature, Humidity and Pressure are done against reference UKAS traceable instruments. The reference system numbers of these instruments are listed above.

All tests have been successfully completed

On behalf of Gill Instruments Ltd

Tony Raine
Quality Control

2002-0396 Issue 1



Gill Instruments Ltd
Saltmarsh Park
67 Gosport Street
Lymington
Hampshire
SO41 9EG, UK

T: +44 (0) 1590 613 500
F: +44 (0) 1590 613 555
E: anem@gill.co.uk

www.gill.co.uk



Copyright © Gill Instruments 2012

Gill Instruments Ltd
Reg No. 3154453 Registered Office: The George Business Centre, Christchurch Road, New Milton, B-25 6QJ



ALS Laboratory Group

ANALYTICAL CHEMISTRY & TESTING SERVICES

SUB-CONTRACTING REPORT

CONTACT	: MR K.W. FAN	WORK ORDER	: HK1327473
CLIENT	: ENVIROTECH SERVICES CO.		
ADDRESS	: SHOP 6, G/F, CASIO MANSION, 209 SHAUKEIWAN ROAD HONG KONG	SUB-BATCH	: 1
		DATE RECEIVED	: 7-OCT-2013
		DATE OF ISSUE	: 22-OCT-2013
PROJECT	: ----	NO. OF SAMPLES	: 1
		CLIENT ORDER	: ----

General Comments

- Sample(s) were received in an ambient condition.
- Sample(s) analysed and reported on an as received basis.
- Calibration was subcontracted to and analysed by Action United Enviro Services.

Signatories

This document has been electronically signed by those names that appear on this report and are the authorised signatories. Electronic signing has been carried out in compliance with procedures specified in the Electronic Transactions Ordinance of Hong Kong, Chapter 553, Section 6.

Signatories

Position

Richard Fung

General Manager

This is the Final Report and supersedes any preliminary report with this batch number.
Results apply to sample(s) as submitted. All pages of this report have been checked and approved for release.

Trading Name: **ALS Technichem (HK) Pty Ltd**

11F, Chung Shun Knitting Centre 1 - 3 Wing Yip Street Kwai Chung N.T. Hong Kong
Tel. +852 2610 1044 Fax: +852 2610 2021 www.alsglobal.com

A Campbell Brothers Limited Company

WORK ORDER : HK1327473
SUB-BATCH : 1
CLIENT : ENVIROTECH SERVICES CO.
PROJECT : ----



ALS Lab ID	Client's Sample ID	Sample Type	Sample Date	External Lab Report No.
HK1327473-001	S/N: 245834	Equipment	07-OCT-2013	S/N: 245834

Equipment Calibration Record

Equipment Calibrated:

Type: Laser Dust monitor
 Manufacturer: Sibata LD-3B
 Serial No. 245834
 Equipment Ref: Nil
 Job Order HK1327473

Standard Equipment:

Standard Equipment: Higher Volume Sampler
 Location & Location ID: AUES office (calibration room)
 Equipment Ref: HVS 018
 Last Calibration Date: 8 October 2013

Equipment Calibration Results:

Calibration Date: 10 & 11 October 2013

Hour	Time	Mean Temp °C	Mean Pressure (hPa)	Concentration in mg/m ³ (Standard Equipment)	Total Count (Calibrated Equipment)	Count/Minute (Total Count/60min)
1hr50min	12:00 ~ 13:50	27.0	1012.8	0.123	6013	56.0
4hr25min	13:55 ~ 18:20	27.0	1012.8	0.207	26441	99.5
16hr20min	18:25 ~ 10:45	27.0	1012.8	0.050	25113	25.7

Sensitivity Adjustment Scale Setting (Before Calibration) 765 (CPM)


Sensitivity Adjustment Scale Setting (After Calibration) 760 (CPM)

Linear Regression of Y or X

Slope (K-factor): 0.0021

Correlation Coefficient 0.9980

Validity of Calibration Record 16 Oct 2013



Operator: Tung Chi Sun Signature:  Date: 16 October 2013

QC Reviewer: Ben Tam Signature:  Date: 16 October 2013

TSP SAMPLER CALIBRATION CALCULATION SPREADSHEET

Location : Gold King Industrial Building, Kwai Chung
 Location ID : Calibration Room

Date of Calibration: 8-Oct-13
 Next Calibration Date: 8-Jan-14

CONDITIONS

Sea Level Pressure (hPa)	1008.1	Corrected Pressure (mm Hg)	756.075
Temperature (°C)	26.8	Temperature (K)	300

CALIBRATION ORIFICE

Make->	TISCH	Qstd Slope ->	2.11662
Model->	5025A	Qstd Intercept ->	-0.01714
Calibration Date->	9-Apr-13	Expiry Date->	9-Apr-14

CALIBRATION

Plate No.	H2O (L) (in)	H2O (R) (in)	H2O (in)	Qstd (m3/min)	I (chart)	IC corrected	LINEAR REGRESSION	
							Slope =	Intercept =
18	3.8	3.8	7.6	1.303	66	65.63	Slope =	21.6637
13	3	3	6.0	1.159	64	63.64	Intercept =	37.9417
10	2.2	2.2	4.4	0.994	60	59.66	Corr. coeff. =	0.9975
8	1.2	1.2	2.4	0.736	54	53.70		
5	0.9	0.9	1.8	0.638	52	51.71		

Calculations :

$$Qstd = 1/m[\text{Sqrt}(H2O(Pa/Pstd)(Tstd/Ta))-b]$$

$$IC = I[\text{Sqrt}(Pa/Pstd)(Tstd/Ta)]$$

Qstd = standard flow rate

IC = corrected chart responses

I = actual chart response

m = calibrator Qstd slope

b = calibrator Qstd intercept

Ta = actual temperature during calibration (deg K

Pstd = actual pressure during calibration (mm Hg

For subsequent calculation of sampler flow:

$$1/m((I)[\text{Sqrt}(298/Tav)(Pav/760)]-b)$$


m = sampler slope

b = sampler intercept

I = chart response

Tav = daily average temperature

Pav = daily average pressure



ALS Technichem (HK) Pty Ltd



ALS Laboratory Group

ANALYTICAL CHEMISTRY & TESTING SERVICES

SUB-CONTRACTING REPORT

CONTACT	: MR K.W. FAN	WORK ORDER	: HK1327471
CLIENT	: ENVIROTECH SERVICES CO.		
ADDRESS	: SHOP 6, G/F. CASIO MANSION, 209 SHAUKEIWAN ROAD HONG KONG	SUB-BATCH	: 1
		DATE RECEIVED	: 7-OCT-2013
		DATE OF ISSUE	: 22-OCT-2013
PROJECT	: ----	NO. OF SAMPLES	: 1
		CLIENT ORDER	: ----

General Comments

- Sample(s) were received in an ambient condition.
- Sample(s) analysed and reported on an as received basis.
- Calibration was subcontracted to and analysed by Action United Enviro Services.

Signatories

This document has been electronically signed by those names that appear on this report and are the authorised signatories. Electronic signing has been carried out in compliance with procedures specified in the Electronic Transactions Ordinance of Hong Kong, Chapter 553, Section 6.

Signatories

Position

Richard Fung General Manager

This is the Final Report and supersedes any preliminary report with this batch number.
Results apply to sample(s) as submitted. All pages of this report have been checked and approved for release.

Trading Name: **ALS Technichem (HK) Pty Ltd**
11/F, Chung Shun Knitting Centre, 1-3 Wing Yip Street, Kwai Chung, N.T., Hong Kong
Tel. +852 2610 1044 Fax. +852 2610 2021 www.alsglobal.com

A Campbell Brothers Limited Company

WORK ORDER : HK1327471
SUB-BATCH : 1
CLIENT : ENVIROTECH SERVICES CO.
PROJECT : ----



ALS Lab ID	Client's Sample ID	Sample Type	Sample Date	External Lab Report No.
HK1327471-001	S/N: IY5546	Equipment	07-OCT-2013	S/N: IY5546

Equipment Calibration Record

Equipment Calibrated:

Type: Laser Dust monitor
 Manufacturer: Sibata LD-3B
 Serial No. 1Y5546
 Equipment Ref: Nil
 Job Order HK1327471

Standard Equipment:

Standard Equipment: Higher Volume Sampler
 Location & Location ID: AUES office (calibration room)
 Equipment Ref: HVS 018
 Last Calibration Date: 8 October 2013

Equipment Calibration Results:

Calibration Date: 10 & 11 October 2013

Hour	Time	Mean Temp °C	Mean Pressure (hPa)	Concentration in mg/m ³ (Standard Equipment)	Total Count (Calibrated Equipment)	Count/Minute (Total Count/60min)
1hr50min	12:00 ~ 13:50	27.0	1012.8	0.123	5911	55.0
4hr25min	13:55 ~ 18:20	27.0	1012.8	0.207	24157	90.9
16hr20min	18:25 ~ 10:45	27.0	1012.8	0.050	24112	24.7

Sensitivity Adjustment Scale Setting (Before Calibration) 629 (CPM)


Sensitivity Adjustment Scale Setting (After Calibration) 633 (CPM)

Linear Regression of Y or X

Slope (K-factor): 0.0023

Correlation Coefficient 0.9992

Validity of Calibration Record 16 Oct 2013



Operator : Tung Chi Sun Signature : *Tung Chi Sun* Date : 16 October 2013

QC Reviewer : Ben Tam Signature : *Ben Tam* Date : 16 October 2013

TSP SAMPLER CALIBRATION CALCULATION SPREADSHEET

Location : Gold King Industrial Building, Kwai Chung
 Location ID : Calibration Room

Date of Calibration: 8-Oct-13
 Next Calibration Date: 8-Jan-14

CONDITIONS

Sea Level Pressure (hPa) 1008.1
 Temperature (°C) 26.8

Corrected Pressure (mm Hg) 756.075
 Temperature (K) 300

CALIBRATION ORIFICE

Make-> TISCH
 Model-> 5025A
 Calibration Date-> 9-Apr-13

Qstd Slope -> 2.11662
 Qstd Intercept -> -0.01714
 Expiry Date-> 9-Apr-14

CALIBRATION

Plate No.	H2O (L) (in)	H2O (R) (in)	H2O (in)	Qstd (m3/min)	I (chart)	IC corrected	LINEAR REGRESSION
18	3.8	3.8	7.6	1.303	66	65.63	Slope = 21.6637 Intercept = 37.9417 Corr. coeff. = 0.9975
13	3	3	6.0	1.159	64	63.64	
10	2.2	2.2	4.4	0.994	60	59.66	
8	1.2	1.2	2.4	0.736	54	53.70	
5	0.9	0.9	1.8	0.638	52	51.71	

Calculations :

$$Qstd = 1/m[\text{Sqrt}(H20(Pa/Pstd)(Tstd/Ta))-b]$$

$$IC = I[\text{Sqrt}(Pa/Pstd)(Tstd/Ta)]$$

Qstd = standard flow rate

IC = corrected chart responses

I = actual chart response

m = calibrator Qstd slope

b = calibrator Qstd intercept

Ta = actual temperature during calibration (deg K

Pstd = actual pressure during calibration (mm Hg

For subsequent calculation of sampler flow:

$$1/m((I)[\text{Sqrt}(298/Tav)(Pav/760)]-b)$$

m = sampler slope

b = sampler intercept

I = chart response

Tav = daily average temperature

Pav = daily average pressure

